

National Radio Research Agency of Korea Equipment Calibration Intervals as of May 14, 2014

Based on RRA Public Notification 2013-6 Chapter 4 (Section 14) on page 9, and in the table on page 16.

The maximum nominal calibration interval shall be as indicated below.

Equipment	Interval (years)
EMI Receiver	1
Antenna	2
Line Impedance Stabilization Network	1
Clamp	1
Coupling/Decoupling Networks (CON)	1
Electric probe	1
Magnetic probe	1
Electrostatic generator	1
Electric Fast Transient(EFI) / Burst Generators	1
Surge generator	1
Voltage dips and variations Testing	1